

Bion Calibrator Data Sheet For Mindray

				Bion Calibrator		LOT		J602				2028.01.30		
				Biorad				1008011/1008012				2029.07.31		
제조사	검사항목	Unit	Method	Bion Cal Conc.	Biorad 1					Biorad 2				
					Mean	SD	Range			Mean	SD	Range		
Bion	ALB	g/dL	BCG	3.7	4.2	0.2	3.8	~	4.6	2.8	0.2	2.4	~	3.2
	ALP	U/L	PNPP/AMP/IFCC	248	143	10	123	~	163	567	25	517	~	617
	AMY	U/L	EPS/IFCC	198	79	4	71	~	87	505	20	465	~	545
	AST(GOT)	U/L	UV without P5P(IFCC)	113	47	4	39	~	55	260	10	240	~	280
	ALT(GPT)	U/L	UV without P5P(IFCC)	119	32	3	26	~	38	111	8	95	~	127
	BUN	mg/dL	Urease GLDH	47.4	16.9	1.5	13.9	~	19.9	47.5	3.5	40.5	~	54.5
	CA	mg/dL	Arsenazo 3	10.9	9.4	0.8	7.8	~	11.0	12.5	1.0	10.5	~	14.5
	CHO	mg/dL	COD/POD	166	244	12	220	~	268	104	6	92	~	116
	CK	U/L	UV Rate	353	129	11	107	~	151	443	20	403	~	483
	CREA	mg/dL	Jaffe	3.82	2.67	0.20	2.27	~	3.07	5.67	0.60	4.47	~	6.87
	D-BIL	mg/dL	Oxidation	2.2	1.81	0.10	1.61	~	2.01	0.43	0.20	0.03	~	0.83
	GGT	U/L	Glupa-C IFCC	100	62	6	50	~	74	171	9	153	~	189
	GLU	mg/dL	Hexokinase	210	83	5	73	~	93	302	15	272	~	332
	GLU 1STEP	mg/dL	GOD/POD	206	85	5	75		95	300	15	270		330
	HDL	mg/dL	Direct	58	91	5	81	~	101	28	3	22	~	34
	IP	mg/dL	Molybdate	5.59	3.55	0.30	2.95	~	4.15	7.57	0.80	5.97	~	9.17
	LDH	U/L	LD(P>L) Tris	573	352	18	316	~	388	863	40	783	~	943
	LDL	mg/dL	Direct	69	110	7	96	~	124	47	4	39	~	55
	T-BIL	mg/dL	Oxidation	4.3	4.13	0.12	3.89	~	4.37	1.12	0.20	0.72	~	1.52
	TG	mg/dL	Enzymetic	146	173	10	153	~	193	99	5	89	~	109
TP	g/dL	Biuret	5.8	6.9	0.35	6.2	~	7.6	4.5	0.2	4.1	~	4.9	
UA	mg/dL	효소법	5.42	4.64	0.35	3.94	~	5.34	10.10	0.7	8.7	~	11.5	

JWBS	CA	mg/dL	Arsenazo 3	10.9	9.4	0.8	7.8	~	11	12.5	1.0	10.5	~	14.5
	CK	U/L	UV Rate	353	129	11	108	~	150	443	20	403	~	483
	IP	mg/dL	Molybdate	5.59	3.55	0.30	2.95	~	4.15	7.57	0.80	5.97	~	9.17
	MG	mg/dL	Xylidyl Blue	2.6	2.0	0.3	1.4	~	2.6	4.3	0.4	3.5	~	5.1

*본 Data는 기종에 따라 오차가 발생할 수도 있음을 알려드립니다.